

# Instructions for Use for Patients NEURO HiTRONIC





Coı	ntent			Page	
1.	Safe	ty Instri	uctions	4	
	1.1	Classif	ication of the Safety Instructions	4	
	1.2	All Ins	tructions for Your Safety	4	
2.	Use			7	
	2.1	Intend	ed Use	7	
	2.2	Indica	tion	7	
	2.3	Contra	indication	7	
	2.4	Qualif	ication	7	
	2.5	Applic	ation	7	
3.	Knee	Joint S	ystem	8	
	3.1		unctions	9	
		3.1.1	Safe Handling of the Joint Functions	9	
		3.1.2	Basic Function in Auto Mode	10	
		3.1.3	Alternative Function in Lock Mode	10	
		3.1.4	Alternative Function in Free Mode	10	
		3.1.5	Alternative Function in Permanent Unlocking	10	
	3.2	Rotary	Switch	11	
	3.3	The Re	emote Control	11	
	3.4	The Us	ser App	12	
		3.4.1	Step Counter	12	
	3.5	The Co	ontroller	12	
	3.6	Manua	al Mode Change	13	
4.	Con	nection	between Controller and Remote Control/App	13	
	4.1	Contro	olling Two Orthoses	13	
5.	Chec	cking th	e Connection between Controller and Remote Control	13	
	5.1	5.1 Indication of the Connection with one Controller			
	5.2	Indica	tion of the Connection with two Controllers	15	
6.	Chec	cking th	e Connection between Controller and User App	15	
7.	Chec	cking th	e Mode and Battery Status	16	
	7.1	Indica	tion of mode and battery status on the controller	16	
	7.2	Indica	tion of Battery Status on the Remote Control/in the App	16	
		7.2.1	Indication of the Battery Status for a Connection with one Controller	17	
		7.2.2	Indication of the Battery Status for a Connection with two Controllers	17	
8.	Ener	gy Cons	umption	18	
	8.1	8.1 Period of Use of Batteries in Different Modes			
	8.2	Energy	r-Saving Modes	18	

9.	Handling of the Battery in the Controller	19
	9.1 Charging the Lithium-Polymer Battery	19
10.	Advice on Using Your Orthosis	19
	10.1 Before Use	19
	10.2 Bluetooth Connection	19
	10.3 The Proper Shoe	19
	10.4 Gait Re-Education	20
	10.4.1 Walking with the Orthosis in Auto Mode	20
	10.5 Malfunction Due to External Impact	21
	10.6 Restrictive Use	21
11.	Maintenance of the Orthosis	21
	11.1 Dirt Removal from the System Joint	22
12.	Storage	22
13.	Advice on Optimal Orthosis Functionality	22
	13.1 System Knee Joint	22
	13.2 Remote control	22
	13.3 Controller	23
14.	Disposal	23
15.	Technical Data	23
	15.1 Ambient Conditions	23
16.	Signs and Symbols	25
17.	CE Conformity	27
18.	Legal Information	27
19.	Electromagnetic Compatibility	28
	19.1 Electromagnetic Environment	28
	19.2 Electromagnetic Emissions for all Devices and Systems	28
	19.3 Electromagnetic Immunity for all Devices and Systems	29
	19.4 Electromagnetic Immunity for Non-Life-Supporting Devices and Systems	30
	19.5 Recommended Safety Distances between Portable and Mobile RF Telecommunication	
	Equipment and the Product NEURO HiTRONIC for Non-Life-Supporting Devices and	
	Systems	31
20.	Handing Over the Orthosis	32

# Instructions for Use for Patients NEURO HiTRONIC System Knee Joint

Dear Patient.

You have received an individually produced orthosis with a high quality FIOR & GENTZ automatic electronic system knee joint from your orthotist or a qualified/trained expert.

# 1. Safety Instructions

# 1.1 Classification of the Safety Instructions

▲ DANGER	Important information about a possible dangerous situation which, if not avoided, leads to death or irreversible injuries.
▲ WARNING	Important information about a possible dangerous situation which, if not avoided, leads to reversible injuries that need medical treatment.
<b>A</b> CAUTION	Important information about a possible dangerous situation which, if not avoided, leads to light injuries that do not need medical treatment.
NOTICE	Important information about a possible situation which, if not avoided, leads to damage of the product.

All serious incidents according to Regulation (EU) 2017/745 which are related to the product have to be reported to the manufacturer and to the competent authority of the Member State in which the orthotist and/or the patient is established.

#### 1.2 All Instructions for Your Safety

# **A** DANGER

#### Potential Traffic Accident Due to Limited Driving Ability

Gather information about all issues concerning safety and security and potential dangers before driving a motor vehicle with orthosis.

# **▲** DANGER

#### Risk of Strangulation Due to Improper Handling of the Cable

Use the orthosis as described in these instructions for use. During use, pay special attention to the connecting cable of the orthosis as well as to the charging cable of the controller.

# **▲** WARNING

#### Jeopardising the Therapy Goal by Not Providing the Necessary Free Movement

Check if the system joint moves freely in order to avoid restrictions of the joint function.

# WARNING

#### Risk of Falling Due to Permanent Higher Load

Do not engage in any sports with the orthosis that expose it to excessive strain. If your patient data has changed (e.g. due to weight gain, growth or increased activity), consult your orthotist or a qualified/trained expert and have them check the suitability of your orthosis with regard to the changed load. You will find the next maintenance appointment in your orthosis service passport.

# **▲** WARNING

#### Risk of Falling Due to Improper Shoe/Wrong Shoe Pitch

Wear a shoe to which your orthosis is adjusted to avoid joint dysfunction.

# **▲** WARNING

#### Risk of Falling Due to Improper Handling

Have your orthotist or a qualified/trained expert inform you about the correct use of the system joint. Disable the permanent unlock function if you no longer wish to use it.

# **A** WARNING

#### Risk of Falling Due to Improper Dirt Removal

In order to avoid a failure of the lock function, remove dirt from the orthosis and the system joint as described in these instructions for use. Do not grease the system joint on your own. If necessary, consult your orthotist or a qualified/trained expert.

# WARNING

#### Risk of Falling Due to Damages to the Orthosis

Avoid damage to your orthosis and to the integrated electronics (e. g. due to shocks, knocks and fall). If your orthosis is still damaged, switch it to Lock mode and contact your orthotist or a qualified/trained expert as soon as possible.

# **A** WARNING

#### Risk of Falling Due to Walking Incorrectly with the Orthosis

Consult your orthotist or a trained/qualified expert about the correct use of your orthosis and the particularities of the system joint. If necessary, we recommend a physiotherapeutic gait re-education.

# **▲** WARNING

#### Risk of Falling Due to Modified Orthosis Settings

If you notice any changes in the orthosis (e.g. loosely attached joint components, loosened screws, play in the system joint or change in performance), immediately contact your orthotist or a qualified/trained expert. Do not secure the screws for the system joint on your own. The adjustments must be checked by your orthotist or a qualified/trained expert before handing over the orthosis and at every maintenance. You will find the next maintenance appointment in your orthosis service passport.

# **▲** WARNING

#### Risk of Injury Due to Improper Handling of the Controller or Remote Control

Use the controller and the remote control as described in these instructions for use. The orthosis may not be worn while it is charging. The controller is a sensitive electronic device with an integrated lithium-polymer battery. When handling the controller please avoid:

- strong heat (e.g. fire, heater, fireplace),
- charging a battery in direct sunlight,
- knocks and shocks (e.g. by pets) as well as
- immersion in water.

# **▲** WARNING

#### Risk of Injury Due to Improper Handling of the System Joint

Use the system joint as described in these instructions for use.

- Do not immerse the system joint in water. The electronic system components (excluding the accessories) are only protected from water splashing on all sides.
- When using the system joint, an opening is formed between the joint's upper and lower parts, in which clothing or skin could get caught.

# NOTICE

#### Limitation of Joint Function Due to Electrostatic/Magnetic Field

Please note that, while using the orthosis, an electromagnetic and magnetic field (e.g. MRI) can lead to ioint dysfunction.

# **NOTICE**

#### Limitation of the Joint Function Due to Lack of Maintenance

Have your orthotist or a qualified/trained expert inform you about the maintenance intervals to be observed in order to avoid joint dysfunctions. You will find the next maintenance appointment in your orthosis service passport.

# NOTICE

#### Damage to Controller Due to Improper Handling

Use the controller as described in these instructions for use. In particular, please ensure that the controller:

- is used with the provided charging cable and power supply unit, and
- is only used at ambient temperatures from -10°C to +40°C.

## NOTICE

#### Damage to Controller and Remote Control Due to Improper Handling

Ensure the correct use in order to avoid joint dysfunction. Avoid:

- to open the controller and the remote control as well as
- using them in areas where radio waves are forbidden (e.g. in planes, hospitals).

Ask the responsible staff on-site about using the controller and remote control.



Any modification of the system joint from your side is not permitted.



In case of problems with the system joint and potentially occurring allergic reactions, contact your orthotist or a qualified/trained expert or the manufacturer. You can find the manufacturer's contact data on the back page of these instructions for use.

#### 2. Use

#### 2.1 Intended Use

The FIOR & GENTZ electrohydraulic system knee joints are exclusively for use for orthotic fittings of the lower extremity. The system joint provides stance phase control and is only allowed to be used for producing a KAFO (knee-ankle-foot orthosis). Every system joint influences the orthosis' function and thus also the function of the leg.

#### 2.2 Indication

The indications for the treatment with an orthosis for the lower extremity are insecurities that lead to a pathological gait. This can be caused, for example, by central, peripheral, spinal or neuromuscular paralyses, structurally conditioned deformities/malfunctions, or surgery.

The physical conditions of the patient, such as muscle strength or activity level are crucial for the orthotic treatment. A safe handling of the orthosis must be ensured. The orthotist or qualified/trained expert selects the appropriate system joints for the orthosis.

#### 2.3 Contraindication

The system joint is not suitable for treatments that were not described in paragraph 2.2, such as a treatment of the upper extremity or a treatment with a prosthesis or ortho-prosthesis, for example after amputations of leg segments.

#### 2.4 Qualification

The system joint must only be handled by an orthotist or a qualified/trained expert.

#### 2.5 Application

All FIOR & GENTZ system joints were developed for everyday life activities such as standing and walking. Extreme loads connected to activities like running, climbing and parachuting are excluded. The system joint can be used at temperatures of  $-10^{\circ}$ C to  $+40^{\circ}$ C.

# 3. Knee Joint System

The knee joint system is equipped with Bluetooth® technology\* and consists of the following components (fig. 1):

- system knee joint
- 2 controller
- 3 remote control for the patient including charging cable with adapter and User app
- 4 Expert app for the orthotist or the qualified/trained expert

The system knee joint and the controller are built into your orthosis. The orthotist or qualified/trained expert uses the Expert app to adjust the orthosis. You need the remote control to operate the orthosis. In addition, you can also use the User app.




Fig. 1

<sup>\*</sup> The Bluetooth® word mark and logos are registered trademarks of Bluetooth SIG, Inc. and any use of these marks by FIOR & GENTZ is under license.

Your orthotist or a trained/qualified expert has provided you with the following system components in addition to your orthosis (fig. 2):



Fig. 2

Item	Art. No.	Description	Unit	Quantity
1	ET0710-01	charging cable	pce.	1
2	PR4000	FIOR & GENTZ lanyard	pce.	1
3	ET3840-P	remote control	pce.	1
4	ET0780	adapter	pce.	1

#### 3.1 Joint Functions

The **NEURO HiTRONIC** is a microprocessor-controlled, automatic system knee joint and provides four joint functions:

- basic function in delivery status in Auto mode
- alternative function in Lock mode
- alternative function in Free mode
- alternative function in permanent unlocking

The essential performance features of the automatic electronic system joint are to remain unlocked in Free mode and locked in Lock mode, as well as to lock or unlock at the right moment in Auto mode.

#### 3.1.1 Safe Handling of the Joint Functions



#### Standing Up From a Seated position

Before standing up, the system knee joint must be set to Free mode. Once you are standing securely, you can switch to Auto or Lock mode.

#### Stance Phase Control

The system joint ensures the infinitely variable stance phase in a knee angle of 0° (5°) to 45°. Stance phase control occurs between 45° and 50° contingent upon the load. If the load limit is exceeded, an overload protection takes effect. It is comparable to an electric fuse. If the overload protection is triggered, this may result in deformation of the roll holder and possibly its fracture. In this case, the roll unit must be replaced by your orthotist or a qualified/trained expert.

#### 3.1.2 Basic Function in Auto Mode

The controller of the orthosis has motion sensors that detect the movement and position of your lower leg. The controller locks and unlocks the system joint depending on which gait phase you are in.

# locked

#### Stance

When you are standing with your orthosis (fig. 3) or when you interrupt the step in stance phase, the NEURO HITRONIC system knee joint locks, as no movement is registered.

Fig. 3

#### Gait

When walking, the system joint locks/unlocks as follows: The system joint is locked in the direction of flexion from mid swing to mid stance. In the gait phases from terminal stance to initial swing, the system joint is unlocked and is therefore free moving (fig. 4).




Fig. 4

(<u>i</u>

If, contrary to expectations, you put weight on the leg with orthosis during the free moving phases, the system joint will not lock.

#### 3.1.3 Alternative Function in Lock Mode

In Lock mode, the system knee joint is a locked joint that prevents a flexion of the leg. An extension remains possible.

#### 3.1.4 Alternative Function in Free Mode

In Free mode, the system knee joint is an unlocked joint that is free moving up to a determined position.

## 3.1.5 Alternative Function in Permanent Unlocking

The system knee joint can be permanently unlocked mechanically, for example for activities such as driving a car or bicycle. In this mode it is guaranteed that the system knee joint will not lock unintentionally.

To do so, take a seat and unlock the system joint manually with the rotary switch below the system joint, by pressing in the rotary switch and then turning it in the direction of the symbol.

If you then press the lock button on the remote control/app, you also save energy. The system knee joint remains unlocked even if you select another mode (e.g. Auto) with the remote control/app. In order to change the system joint's mode again with the remote control/app, turn the rotary switch in the direction of the symbol (fig. 5).

Fig. 5

# 3.2 Rotary Switch

There is a rotary switch below the system joint (fig. 5). The three symbols show what happens if the rotary switch is turned in their respective direction. The arrows below show how the rotary switch should be controlled in order to select a symbol.

Symbol	Action	Meaning		
Turn the rotary switch in the centre and press in  Press in the rotary switch and turn it in the direction of the		The system joint is operated from the remote control/app.		
		The system joint is unlocked as long as the rotary switch is pressed in. As soon as it is released, the system joint will be operated once again from the remote control/app.		
		The system joint is now in permanent unlocking (see paragraph 3.1.5).		

#### 3.3 The Remote Control

You can select the mode on your orthosis with the remote control. Make sure that you are standing securely when changing the mode of your orthosis. Every time you press a button on the remote control, the LED flashes briefly.

Remote Control	Item	Description	Meaning
Auto o	1	LED	The LED indicates light signals for the selected mode and the battery health.
Free 3	2	Auto button	The system joint switches into Auto mode.
Lock 4	3	Free button	The system joint switches into Free mode.
ACC & CLANE	4	Lock button	The system joint switches into Lock mode.

Handle the remote control properly. If your remote control does not work as usual, do not try to open it. Consult your orthotist or a qualified/trained expert.

# 3.4 The User App

The app is intended to complement your remote control. It offers the same range of functions.





Fig. 6



The orthosis can only be operated with remote control or app to which it is currently connected. Other remote controls/apps have no influence on your orthosis.

\* Apple Watch is a trademark of Apple Inc., registered in the U.S. and other countries.

# 3.4.1 Step Counter

The app gives you access to the step counter, which counts all steps you take with the leg with orthosis in the different modes. If you would like to know how many steps you have taken in total (with both legs), double the value.



Fig. 7

#### 3.5 The Controller

The controller is mounted to your orthosis. It receives commands from the remote control/app, registers your movements and controls the system knee joint.

Controller with Integrated Lithium-Polymer Battery	Item	Description
	1	multicolour LED display for battery charging, mode and Bluetooth connection
1	2	MODE button
E C	3	charging connection
2		

## 3.6 Manual Mode Change

A MODE button, through which the orthosis can be operated manually, is built into the controller.

Depending on the mode that is already selected, it can be set in the following order: Auto, Free and Lock. This button is particularly important if you are travelling by plane, as it is possible that you may not be allowed to use the remote control/app during take-off, final approach or landing. You are allowed to use the remote control/app during the flight and after landing.



The MODE button can only be used as long as the battery is not fully discharged. When the battery is fully discharged, you can only use the Lock mode.

# 4. Connection between Controller and Remote Control/App

The connection between the controller and the remote control is established by your orthotist or a qualified/ trained expert. If you wish to operate the controller with the User app, use the app menu and select the desired menu item to establish a connection. Follow the additional instructions in the app.

# 4.1 Controlling Two Orthoses

If you are wearing two orthoses with a **NEURO HITRONIC** knee joint system, you have the option of connecting the controllers of both orthoses with one or two remote controls. If you activate two remote controls, you can change the modes separately for each controller/orthosis. If you activate only one remote control, the modes of both controllers/orthoses are changed simultaneously. You can change the modes for both controllers or orthoses separately or at the same time, with the **User** app.

# 5. Checking the Connection between Controller and Remote Control

Due to signals of the controller and the remote control, you will be informed if your remote control is connected with the controller. The LED at the remote control indicates that the remote control and the controller communicate with each other. There are different signals for a connection with one or two controllers.



If you have previously operated the orthosis with the app, you have to close the app in order to operate the orthosis with the remote control.

# 5.1 Indication of the Connection with one Controller

Remote Control	Light Signal	Meaning		
	colour: yellow, green, red (depending on battery status)			
	signal duration: ■	The remote control is connected with the controller. The command		
Auto	Auto	with the controller. The command was sent successfully.		
Lock	colour: red	- The orthosis is in Sleep mode (see paragraph 8.2).		
2000 ( 0000)	signal duration: ■ ■			
One of the three buttons was pressed.	Auto	<ul><li>The battery is empty.</li><li>The remote control is too far from the orthosis.</li></ul>		
	colour: blue (controller)			
	signal duration: ■	The controller communicates with		
	Auto	the remote control.		

#### 5.2 Indication of the Connection with two Controllers

Remote Control	Light Signal	Meaning		
	colour: yellow, green, red (depending on battery status)			
	signal duration: ■	The remote control is connected with the controllers. The com-		
		mand was sent successfully to an orthosis.		
	colour: red	- The orthoses are in Sleep mode		
	signal duration: ■ ■	(see paragraph 8.2).		
Auto	Auto	- The batteries are empty The remote control is too far from the orthoses.		
Lock	colour: yellow, green, red (depending on battery status) and afterwards red	The command was sent successfully to an orthosis		
One of the three buttons was pressed.	signal duration: ■ ■ ■	- The second orthosis is in Sleep mode (see paragraph 8.2) The battery for the second ortho-		
·	Auto	sis is empty.  - The remote control is too far from the second orthosis.		
	colour: blue (controller)			
	signal duration: ■	The controllers communicate with		
	Auto	the remote control.		

# 6. Checking the Connection between Controller and User App

In order to operate the orthosis via the app, Bluetooth must be permanently switched on and the app must be open in the foreground. Use the menu of the app and select the desired menu item to connect to one or two controllers. Follow the additional instructions in the app.

An orthosis can only be operated with one app at the same time because there is a connection between the controller and the app. Other apps have no influence on the connected orthosis. You can continue to use the remote control instead of the app when the controller is connected to the remote control and is not actively communicating with the app. If there is an active connection to the app, the blue LED on the controller blinks permanently and the controller cannot be operated with the remote control. The app is intended to complement your remote control.

# 7. Checking the Mode and Battery Status


#### 7.1 Indication of mode and battery status on the controller

You can see the mode and battery status of the controller on the remote control or in the app. Furthermore, the LED battery level indicator displays the following light signals for the battery status:

Light Signal	Meaning	
colour: yellow, green, red (depending on battery status)	The controller is in Auto mode.	
signal duration: ■		
colour: yellow, green, red (depending on battery status)	The controller is in Free mode.	
signal duration: ■■		
-	The controller is in Lock mode.	

The battery status is not displayed in Lock mode. It can be viewed on the remote control or in the app.

The controller emits the following sounds for battery status when the battery is almost empty:



Due to the importance of a properly functioning orthosis, this signal sounds every minute. This period can be extended to ten minutes by pressing one of the three mode buttons on the remote control/app. To do this, select the mode in which your orthosis is currently operating so that you do not inadvertently change modes. After ten minutes, the pause can always be extended by another ten minutes by pressing the mode button again. If no mode button is pressed, the signal will sound every minute. The sound signals for the battery status can be turned off until the next charge through the User app settings.

# 7.2 Indication of Battery Status on the Remote Control/in the App

If you operate the orthosis via the app, you can see the battery status of the controller(s) at any time in the app.

You can also check the status of the batteries in the controller(s) with your remote control. There are different signals for a connection with one or two controllers.

16

7.2.1 Indication of the Battery Status for a Connection with one Controller

Light Signals for the Remote control (with Auto Mode Example):

	Light Signal				
Remote Control	LED	Colour	Signal Duration	Meanin	g
	Auto	green			The battery for the controller is fully charged.
Audo Free Lock SHEDS-SHENSE	Auto	yellow	•		The battery status is low. Depending on the battery health, it takes max. 7 hours to fully discharge.
One of the three buttons was pressed.	Auto	red	•		The battery is almost empty. Depending on the battery health, it takes max. 3 hours to fully discharge.

# 7.2.2 Indication of the Battery Status for a Connection with two Controllers

If your remote control is connected with two controllers, the light signal on the remote control does not apply automatically to both controllers but to that with the lowest battery. The battery status indicator on the controller or in the app (see paragraph 7.1) allows you to determine whether the batteries of both controllers are affected, or which orthosis requires its battery to be charged.

Example: The LED of the remote control flashes red, after you have pressed one of the three buttons. The LED battery status indicator of the controller for the right orthosis blinks green, which means that the battery is full. The LED battery status indicator of the controller for the left orthosis blinks red. This orthosis should soon be charged.

# 8. Energy Consumption

#### 8.1 Period of Use of Batteries in Different Modes


The following average battery life was determined at room temperature:

Auto	Free	Lock
26 000 strides	24 hours	over 2 weeks

#### 8.2 Energy-Saving Modes

Your orthosis has three different energy-saving modes:

- If the orthosis is not moved in Auto or Free mode for more than two hours, the orthosis switches into
  Lock mode automatically. The Lock mode saves energy. If you press any button on the remote control, the
  orthosis will change back from Lock mode into the desired mode.
- If the orthosis is not moved for more than 30 minutes in Lock mode, it will automatically switch into Sleep mode. In Sleep mode, the orthosis consumes very little energy. The controller no longer receives signals from the remote control/app. To switch back into Lock mode, move the orthosis slightly. The LED on the controller then lights up briefly in all colours.



If the orthosis is not moved for more than 72 hours, it switches automatically into Deep Sleep mode. In
Deep Sleep mode the controller does not consume energy and no longer receives signals from the remote
control/app. In order to put the orthosis back in operation, press the MODE button on the controller or
connect the charging cable.

# 9. Handling of the Battery in the Controller

The controller has a long service life and battery lifespan. Do not try to disassemble the controller as the battery is a fixed part of the controller.

## 9.1 Charging the Lithium-Polymer Battery

You can charge the battery using the charging cable and power supply unit included in the scope of delivery, via a common household power socket. Always charge the battery fully and respect the general conditions of use and storage.

If the period of use of the orthosis is considerably shortened despite the fully charged battery, contact your orthotist or a qualified/trained expert.

# 10. Advice on Using Your Orthosis

#### 10.1 Before Use

Pay attention to the following every time you use the orthosis:

- Check the battery status of the controller.
- Set the orthosis in Free mode in order to put it on.
- Set the rotary switch to the symbol so that you can operate the system joint via the remote control/app.

#### 10.2 Bluetooth Connection

The connection quality depends on how interference-free your environment is.

#### 10.3 The Proper Shoe

The orthotist or qualified/trained expert adjusts your orthosis when you try it on and make your first steps. Since you have to get used to your new orthosis, the adjusted orthosis should be regularly checked in the first weeks and, if necessary, newly adapted to your need for safety. The orthosis is adjusted to the pair of shoes (shoe pitch) with which you start walking with your orthosis. If you want to wear other shoes, your orthotist or a qualified/trained expert must ensure that the orthosis is also adjusted to these shoes.

#### 10.4 Gait Re-Education

In order to be able to use your orthosis optimally, you should make use of a physiotherapeutic gait re-education.

In gait re-education, the following should be specially trained:

- walking upright, with the upper body slightly bent forward;
- applying as little body weight as possible on walking assistive devices (such as canes, parallel bars or walkers) as otherwise the physiological gait can be affected.

With gait re-education, you become more secure in using your orthosis, your gait pattern improves and you get used to your orthosis sooner. This is particularly important when a locked orthosis has been used before for many years. Gait re-education can also be supported by acoustic signals from the orthosis. Ask your orthotist or a qualified/trained expert for additional information. Be extremely careful when managing stairs, going on uneven surfaces or uphill/downhill. If you do not feel secure in using your orthosis in Auto mode yet, we recommend selecting Lock mode (see paragraph 3.1.3). Regularly report your experiences with the orthosis during the first weeks to your orthotist or a qualified/trained expert . This is the only way you can get specific advice or help.

.....



The more physiological the gait characteristics become, the better the orthosis can support you.

## 10.4.1 Walking with the Orthosis in Auto Mode

The system knee joint unlocks during the first stride. The e-motion algorithm makes it possible for the motion sensors of the controller to unlock the system knee joint at the right moment, when the second step is taken. Therefore, it is not necessary to become accustomed to always taking the first step using the leg with the orthosis. This leads to a more physiological gait pattern, especially when both legs are treated.

# 10.5 Malfunction Due to External Impact

The system knee joint is provided with electronic components that react sensitively to very strong shocks. This can cause the system joint not to remain unlocked during the swing phase but to lock. The orthosis should then work in the previously set mode again. If this is not the case, change the mode with your remote control/app.

In general, try to avoid great damages to your orthosis, e.g due to shocks, knocks or falls because this may lead to failure of different system components and, in the worst case, of the orthosis. If you notice damage on the orthosis, use it exclusively in Lock mode and contact your orthotist or a qualified/trained expert as soon as possible.



If there is a failure of the joint function, the orthosis switches automatically into Lock mode. Thus, it enables stability during stance phase and reduces the risk of falling.



If you would like to completely turn off the orthosis for safety reasons, press and hold the MODE button for approx. 17 seconds. This will cause a short beep to sound. After 6–10 seconds a longer beep will sound, and after an additional 10 seconds an extra long beep will sound. The orthosis will then switch into Deep Sleep mode (complete power interruption). If you would like to use the orthosis again, turn it back on by pressing the MODE button or by plugging in the charging cable.

#### 10.6 Restrictive Use

The system knee joint has been checked for its electromagnetic compatibility by European law. That means the orthosis works in an electromagnetic environment without introducing intolerable electromagnetic disturbances to other devices in that environment. Nevertheless, similar to smartphones, pay attention to whether or under what conditions you are allowed to use your orthosis in specially designated areas, because the integrated electronics use radio waves (Bluetooth) and can be affected by these as well. In specially designated areas (fig. 8), ask the responsible staff on-site if you can use the orthosis without restrictions. If you are not allowed to use the remote control/app, change the mode using the MODE button (see paragraph 3.6) or, if needed, take off the orthosis.



Fig. 8

#### 11. Maintenance of the Orthosis

Have your orthotist or a qualified/trained expert check the system joint of your orthosis during regular maintenance. When the orthosis is handed over to you, you receive an orthosis service passport. Bring it to each follow-up and have your orthotist or a qualified/trained expert enter the next maintenance appointment. For your own safety, respect the maintenance appointments. Never carry out maintenance work or other adjustments and repairs yourself. In the case of children and people with cognitive impairments, we would like to point out to you as parents or care team that you must regularly check the orthosis and the system joint for signs of wear. If you notice any changes, immediately contact your orthotist or a qualified/trained expert.

#### 11.1 Dirt Removal from the System Joint

Remove dirt from the system joint on a regular basis. Use a dry cloth and clean the system joint only superficially. Then, remove visible dust and lint from the mechanics by using tweezers. Check the orthosis in straight and flexed position.

# 12. Storage

We recommend that you do not store the system joint in a damp environment.

# 13. Advice on Optimal Orthosis Functionality

If you are using the User app, you can display a troubleshooting code if problems occur with your orthosis. You can then send this code to your orthotist or a qualified/trained expert so that the error can be corrected more quickly. You can find the code for troubleshooting in the app under the menu item "Information".

#### 13.1 System Knee Joint

Problem	Cause	Action
The system joint switches unintentionally into Lock mode	The battery is empty.	Charge the battery.
The system joint switches unintentionally into permanently unlocked mode	There is a problem with the electronics.	Press and hold the MODE key for 17 seconds. This will cause a short beep to sound. After 6–10 seconds a longer beep will sound, and after an additional 10 seconds an extra long beep will sound. The orthosis will then switch into Deep Sleep mode (total power interruption) and will remain in Lock mode. The orthosis can still be used in Lock mode. Consult your orthotist or a qualified/trained expert.
The system joint remains unlocked.	The rotary switch is unlocked.	Set the rotary switch to the symbol.

#### 13.2 Remote control

Problem	Cause	Further Action
The controller does not respond	There is an active connection between the User app and the controller.	Check whether the controller is still connected to the User app. Close the app.
to pressed buttons on the remote control.	You are moving while pressing the button.	Remain still while pressing a button.
	The controller is in Sleep mode.	Move the orthosis slightly.

# 13.3 Controller

Problem	Cause	Further Action
When the MODE button is pressed, the LEDs do not light up.	The battery is not charged.	Charge the battery. If the problem remains, contact your orthotist or a qualified/trained expert.
No devices are found during connection of the controller and the User app.	The controller was not in connecting mode.	Within 30 seconds of pressing the MODE button, establish a connection between the User app and controller (see paragraph 4). Check whether the LEDs light up (see paragraph 5.1) or whether a short and a longer beep tone can be heard. If the problem remains, contact your orthotist or a qualified/trained expert.

# 14. Disposal

If you no longer need the orthosis, please return it to your orthotist or a qualified/trained expert. The product must not be disposed of with the residual waste (fig. 9). If you have a defective controller, please also return it to your orthotist or a qualified/trained expert.



Fig. 9

# 15. Technical Data

NEURO HITRONIC	
period of use	unlimited, excluding wear parts
protection type	IP44
operating mode	continuous operation

# 15.1 Ambient Conditions

Operation	
ambient temperature	-10°C - +40°C
	+5°C - +40°C when charging the battery, no exposure to direct sunlight
relative air humidity	0% – 95%, non-condensing air humidity
air pressure	1060mbar – 700mbar

Transport	
ambient temperature	-25°C - +60°C
relative air humidity	without original packing: max. 95%, non-condensing air humidity with original packing: max. 95%
air pressure	1060mbar – 700mbar

Storage	
ambient temperature	+5°C - +40°C, no exposure to direct sunlight
relative air humidity	max. 95%, non-condensing air humidity
air pressure	1060mbar – 700mbar

Data Transmission	
remote technology	bluetooth low energy
working range	min. 2m
frequency range	2402MHz – 2480MHz
modulation	GFSK
data rate (OTA)	1Mbps
maximum output power (EIRP)	+5dBm

Power Supply Unit with Charging Cable (not Part of the Medical Device)			
article number	ET0780		
manufacturer's designation	HNP12-USBV2, HNP07-USBV2		
ambient temperature in operation	-10°C - +40°C		
ambient temperature in storage	-20°C - +70°C		
relative air humidity	10% – 90%rH		
input voltage	90V – 264V (AC)		
input frequency	47Hz – 63Hz		
power	12W		
output voltage	5V (DC)		
output current	max. 2.4A		
Charging Cable (not Part of the Medical Device)			
article number	ET0710-01		

Controller Battery		
type	lithium-polymer battery	
operating time at room temperature and full battery charge after 3 years of use	Auto mode: 26 000 strides/Free mode: 24 hours	
behaviour of the system knee joint during the charging process	The system knee joint has no function.	

User and Expert App	
supported operating system	at least Android 6.0 or iOS 10

# 16. Signs and Symbols

 $\epsilon$ 

CE labelling according to Regulation (EC) 2017/745 for medical devices



medical device



catalogue number



Do not dispose of electronic devices with household waste. Dispose of the device and accessories at official delivery points for electronic devices.



manufacturer



batch code



protect from heat



keep dry



temperature limit values for storage/for transportation



air humidity limit values for storage/for transportation



air pressure limit values for storage/for transportation



follow the instructions for use



single patient - multiple uses

IP44

protection from the ingress of solid foreign bodies (diameter  $\geq$  1.0 mm) and from splashing water on all sides



Unique Device Identifier – product identification number

#### Remote Control Type Plate



# **Controller Type Plate**





www.fior-gentz.de ABCDEF-UVXYZ ET3850

26

# 17. CE Conformity

We declare that our medical devices as well as our accessories for medical devices are in conformity with the requirements of Regulation (EU) 2017/745. Therefore, the FIOR & GENTZ products bear the CE marking.

The product satisfies the requirements of the RoHS Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011, for limiting the use of specific hazardous substances in electrical and electronic equipment.

# 18. Legal Information

With the purchase of this product, our General Terms and Conditions of Business Transactions, Sales, Delivery and Payment will apply. The manufacturer shall bear liability when the product is used according to the specifications in the instructions for use. The manufacturer shall not be liable for damages caused from non-observance of these instructions for use, especially from improper use or unauthorised alteration of the product. The warranty expires, for example, if the product is mounted several times. Please note that the product is not supposed to be combined with other components or materials than with those recommended according to the configuration result from the FIOR & GENTZ Orthosis Configurator. Should the product be combined with products from other manufacturers, a written consent by the seller is required.

The information in the instructions for use is valid at the time of printing. The contained product information serves as guidelines. Subject to technical modifications.

All rights, particularly the distribution, copy and translation of this manual or any part of it, in paper or as electronic document, must be authorised in writing by FIOR & GENTZ Gesellschaft für Entwicklung und Vertrieb von orthopädietechnischen Systemen mbH. Reprints, copies and any other electronic reproduction, even partial, must be authorised in writing by FIOR & GENTZ Gesellschaft für Entwicklung und Vertrieb von orthopädietechnischen Systemen mbH.

# 19. Electromagnetic Compatibility

Special precautions must be taken for all electronic medical devices as regards electromagnetic compatibility (EMC). This device complies with standard IEC 60601-1-2:2014.

- All electronic medical devices must be installed and put into operation in compliance with the EMC-relevant information contained in these instructions for use.
- Portable and mobile RF communication devices may interfere with the performance of electronic medical devices.

The device satisfies all valid and required standards for electromagnetic disturbances.

- It generally has no effect on systems and devices found in its vicinity.
- It is generally not affected by systems and devices found in its vicinity.
- It is not safe to operate the device in the vicinity of high-frequency surgical devices.
- It is recommended that the device not be used in the direct vicinity of other devices.

#### 19.1 Electromagnetic Environment

Operation of the device is allowed in the following electromagnetic environments:

- professional health care facilities (e.g. hospital, etc.)
- home health care areas (e.g. use at home, use outdoors)

The patient must ensure that the device is exclusively operated in such environments.

#### 19.2 Electromagnetic Emissions for all Devices and Systems

Usage Instructions and Manufacturer's Declaration - Electromagnetic Emissions

The product NEURO HiTRONIC is designed for operation in an electromagnetic environment as specified below. The customer or user of the product NEURO HiTRONIC must ensure that it is operated exclusively in such an environment

Interference Measurements	Compliance	Usage Instructions for Electromagnetic Environment
RF emissions according to CISPR 11	group 1	The product <b>NEURO HiTRONIC</b> uses RF energy only for its internal function. Therefore, the RF emissions are very low and unlikely to cause any interference in nearby electronic devices.
RF emissions according to CISPR 11	class B	The product NEURO HiTRONIC is suitable for use
harmonics according to IEC 61000-3-2	class A	outside of residential facilities. It is also suitable for facilities directly connected to a public low-voltage
voltage fluctuations/flicker according to IEC 61000-3-3	complies with requirements	network that supplies residential buildings.

Usage Instructions and Manufacturer's Declaration - Electromagnetic Immunity

19.3 Electromagnetic Immunity for all Devices and Systems

The product NEURO HiTRONIC is designed for operation in an electromagnetic environment as specified below. The customer or user of the product NEURO HiTRONIC must ensure that it is operated exclusively in such an environment.

Immunity Test	Test Level IEC 60601	Compliance Level	Usage Instructions for Electromagnetic Environment
electrostatic discharge (ESD) according to IEC 61000-4-2	± 8kV discharge on contact ± 2kV, ± 4kV, ± 8kV, ± 15kV discharge through air	± 8kV discharge on contact ± 15kV discharge through air	Floors should be made of wood or concrete or be ceramic tiled. If the floor covering is made of synthetic material, the relative humidity must be at least 30%.
electrical fast tran- sients/bursts according to IEC 61000-4-4	100kHz pulse renetial supply lines be equivalent to t		The quality of the supply voltage should be equivalent to that of a typical commercial or hospital environment.
surges according to IEC 61000-4-5	± 0.5kV, ± 1kV line- to-line voltage ± 0.5kV, ± 1kV line- to-ground voltage	± 1kV line-to- line voltage ± 1kV line-to- ground voltage	The quality of the supply voltage should be equivalent to that of a typical commercial or hospital environment.
voltage drops, short interruptions and fluc- tuations of the supply voltage according to IEC 61000-4-11	0% of $\rm U_T$ for 0.5 cycles and phase angles of 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315° 70% of $\rm U_T$ for 25/30 cycles and phase angles of 0° 0% of $\rm U_T$ for 250/300 cycles	0% of U <sub>T</sub> for 0.5 cycles and phase angles of 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315° 70% of U <sub>T</sub> for 25/30 cycles and phase angles of 0° 0% of U <sub>T</sub> for 250/300 cycles	The quality of the supply voltage should be equivalent to that of a typical com- mercial or hospital environment.
magnetic field at mains frequency (50, 60Hz) according to IEC 61000-4-8	30A/m	30A/m	The magnetic fields at mains frequency should be equivalent to the typical levels of a commercial or hospital environment.

29

# 19.4 Electromagnetic Immunity for Non-Life-Supporting Devices and Systems

Usage Instructions and Manufacturer's Declaration - Electromagnetic Immunity

The product **NEURO HiTRONIC** is designed for operation in an electromagnetic environment as specified below. The customer or user of the product **NEURO HITRONIC** must ensure that it is operated exclusively in such an environment.

Immunity Test	Test Level IEC 60601	Compliance Level	Usage Instructions for Electromagnetic Environment	
conducted RF inter- ference according to IEC 61000-4-6	3V <sub>rms</sub> 150kHz to 80MHz 6V <sub>rms</sub> in ISM bands 150kHz to 80MHz	$3V_{rms}$ 150kHz to 80MHz 6 $V_{rms}$ in ISM bands 150kHz to 80MHz	Portable and mobile wireless devices should be used at a safety distance from the product NEURO HiTRONIC and its lines. The recommended safety distance was calculated using the equation applicable to the transmission frequency.	
radiated RF interfer- ence according to IEC 61000-4-3	10V/m 80MHz to 2.7GHz 80% AM 1kHz	10V/m 80MHz to 2.7GHz	Recommended safety distance: d = 1.2 √P d = 1.2 √P 80MHz to 800MHz d = 2.3 √P 800MHz to 2.7GHz P is the nominal output of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended safety distance in metres (m). According to an on-site investigation³, the field strength of stationary radio transmitters should be below the compliance level at all frequencies.  Interference may occur in the vicinity of devices marked with the following symbol:	

Note 1: The higher frequency range applies between 80MHz and 800MHz.

Note 2: These guidelines may not be applicable in all cases. The propagation of electromagnetic factors is affected by absorption and reflection from buildings, objects and people.

<sup>&</sup>lt;sup>a</sup> The field strength of stationary RF transmitters such as base stations of radio telephones and mobile land radio equipment, amateur radio stations, AM and FM radio and television stations cannot be precisely determined in advance. A site survey is recommended to establish the electromagnetic environment as a result of stationary RF transmitters. If the field strength determined at the site of the product NEURO HiTRONIC exceeds the compliance level specified above, the product NEURO HiTRONIC has to be monitored with regard to normal operation during use. If unusual performance characteristics are noted, additional measures may be necessary, such as changing the orientation or site of the product NEURO HiTRONIC.

# 19.5 Recommended Safety Distances between Portable and Mobile RF Telecommunication Equipment and the Product NEURO HiTRONIC for Non-Life-Supporting Devices and Systems

Usage Instructions and Manufacturer's Declaration – Recommended Safety Distances between Portable and Mobile RF Telecommunication Equipment and the Product NEURO HiTRONIC

The product NEURO HiTRONIC is designed for operation in an electromagnetic environment where RF interference is monitored. The customer or user of the product NEURO HiTRONIC can help prevent electromagnetic interference by complying with the minimum distances between portable and mobile RF communication equipment (transmitters) and the product NEURO HiTRONIC, as specified below according to the maximum output of the communication equipment.

Nominal Output of the Transmitter [W]	Safety Distance [m] According to Transmission Frequency		
	150kHz to 80MHz d = 1.2 √P	80MHz to 800MHz d = 1.2 √P	800MHz to 2.5GHz d = 2.3 √P
0.01	0.12	0.12	0.23
0.1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

For transmitters whose maximum nominal output is not specified in the table above, the recommended safety distance d in metres (m) can be determined using the equation in the respective column, where P stands for the maximum nominal output of the transmitter in watts (W) according to the transmitter manufacturer.

Note 1: The higher frequency range applies between 80MHz and 800MHz.

Note 2: These guidelines may not be applicable in all cases. The propagation of electromagnetic factors is affected by absorption and reflection from buildings, objects and people.

# 20. Handing Over the Orthosis

When handing over the orthosis to the patient, parents or care team by the orthotist or qualified/trained expert, they also received the instructions for use for patients as well as the orthosis service passport. The functions and handling of the orthosis were explained in detail by means of these instructions for use. Ente the next maintenance appointment in the orthosis service passport.			
Place, Date	Signature Orthotist or Qualified/Trained Expert		



# **ORTHOSIS SERVICE PASSPORT**

Have you not yet received an orthosis service passport? Ask your orthotist or a qualified/trained expert!







