



ichroma™ COVID-19 Ab

INTENDED USE

ichroma™ COVID-19 Ab is a fluorescence Immunoassay (FIA) for the qualitative determination of IgG/IgM antibodies against 'Novel Coronavirus' in human whole blood /serum/ plasma. It is helpful as an aid in the screening of early mild, asymptomatic or acute patients for identification of 'Novel Coronavirus (eg, SARS-CoV-2)' infection with high sensitivity.

For *in vitro* diagnostic use only.

INTRODUCTION

The third zoonotic human coronavirus (CoV) of the century emerged in December 2019, with a cluster of patients connected to Wuhan, Hubei Province, China. This virus, the newly identified coronavirus 2019 nCoV, could cause risky pneumonia so that prevention and control of the infection has become highly required. The 2019 -nCoV is a member of the Betacoronavirus Genus, that also includes Severe Acute Respiratory Syndrome coronavirus (SARS-CoV) and Middle East Respiratory Syndrome coronavirus (MERS-CoV). Since it is identified that symptoms become rapidly severe without a proper treatment after onset of illness, early diagnosis of the virus infection is quite crucial. Currently, the spread of the viral transmission become fast so that the prevention of local transmission requires a point-of care test (POCT), which shows quick outcome within 20 minutes.

ichroma™ COVID-19 Ab test is an *in vitro* diagnostic medical device that helps you to diagnose Novel Coronavirus infections quickly and accurately by measuring the IgG or IgM antibody for the 2019-nCoV.

* The benefits of using this product are;

- 1) To prevent the spread (secondary infection) and recovery of CoV infections, the most important serological test results, determined between the first two weeks after infection, can increase the confidence of confirmatory testing with RT-PCR.
- 2) Periodic serological tests after an infection is confirmed can help determine when to end treatment by analyzing the formation of protective antibodies through seroconversion and recovery of infection through treatment.

PRINCIPLE

This test uses a sandwich immunodetection method; fluorescence-labeled conjugates in a dried detection buffer (DB) binds to antibody in sample, forming antibody-antigen complexes, and migrates onto nitrocellulose matrix to be captured by the other immobilized-anti-human IgG & anti-human IgM on test strip. More antibodies in sample forms the more antigen-antibody complexes which lead to stronger fluorescence signal by detector antigen, which is processed by to instrument for ichroma™ tests to show anti-COVID-19 IgG and IgM concentration of in the sample respectively.

COMPONENTS

ichroma™ COVID-19 consists of 'cartridges', 'detectors', 'detector diluent', 'ID chip' an 'Instruction for use'.

- The cartridge part contains the membrane called a test strip which has anti-human IgM at the test line 1, anti-human IgG at the test line 2 and chicken IgY at the control line. All cartridges are packed in a box with an ID chip and each cartridge is individually sealed in an aluminum foil pouch containing a desiccant.
- The detector has a granule containing the viral antigen-fluorescence conjugate, anti-chicken IgY-fluorescence conjugate, bovine serum albumin (BSA) as a stabilizer and sodium azide as a preservative. All detectors are packed in a pouch.
- The detector diluent contains salt, detergent and sodium azide as a preservative in Tris buffer and it is pre-dispensed in a vial. The detector diluent is packed in a box.

WARNINGS AND PRECAUTIONS

- For *in vitro* diagnostic use only.
- Follow the instructions and procedures described in this 'Instruction for use'.
- Do not reuse cartridges or detector. A cartridge should be used for testing on sample only. A detector should be used for processing of one sample only.
- Use only fresh samples and avoid direct sunlight.
- Lot numbers of all the test components (Cartridge, detector and ID chip) must match each other.
- Do not interchange the test components between different lots or use the test components after the expiration date, either of which might yield incorrect test result(s).
- The cartridge should remain sealed in its original pouch just before use. Do not use the cartridge if pouch is damaged or has already been opened.
- Frozen sample should be thawed only once. For shipping, samples must be packed in accordance with local regulations. Sample with severe hemolysis and/or hyperlipidemia must not be used.
- Allow the cartridge, detector, detector diluent and sample to be at room temperature for approximately 30 minutes

before use.

- The instrument for ichroma™ tests may generate slight vibration during use.
- Do not eat the detector and detector diluent. Any components intake could cause diarrhea or vomiting.
- The detector diluent contains Na₃N as preservatives, of which the contact to eyes, skin or clothing should be avoided. If it happens, please wash with running water immediately.
- Please apply the sample and detector mixture exactly for accurate test result. Or it may cause erroneous results.
- The ichroma™ instruments may generate slight vibration during use.
- Used cartridges, detector tubes and pipette tips should be handled carefully and discarded by an appropriate measure in accordance with the relevant local regulations.
- An exposure to larger quantities of sodium azide may cause specific health issues like convulsions, low blood pressure and heart rate, loss of consciousness, lung injury and respiratory failure.
- **ichroma™ COVID-19 Ab** will provide accurate and reliable results subject to the below conditions.

- **ichroma™ COVID-19 Ab** should be used only in conjunction with the instrument for ichroma™ tests.

- **Have to use recommended anticoagulant sample.**

Recommended anticoagulant

Na EDTA, K₂ EDTA, Na-Heparin,

Li-heparin, Sodium citrate

STORAGE AND STABILITY

| Storage condition | | | |
|-------------------|---------------------|------------|------------|
| Component | Storage Temperature | Shelf life | Note |
| Cartridge | 2 - 30 °C | 20 months | Disposable |
| Detector tube | 2 - 30 °C | 20 months | Disposable |
| Detector diluent | 2 - 30 °C | 20 months | Unopened |
| | | 12 months | Opened |

- After the cartridge pouch is opened, the test should be performed immediately.

LIMITATION OF THE TEST SYSTEM

- The test may yield false positive result(s) due to the cross-reactions and/or non-specific adhesion of certain sample components to the capture/detector antibodies.

- The test may yield false negative result(s) due to the non-responsiveness of the antigen to the antibodies which is most common if the epitope is masked by some unknown components, so therefore not being able to be detected or captured by the antibodies. The instability or degradation of the antigen with time and/or temperature may also cause the false negative result as it makes antigen unrecognizable by the antibodies.
- Other factors may interfere with the test and cause erroneous results, such as technical/procedural errors, degradation of the test components/reagents or presence of interfering substances in the test samples.
- Any clinical diagnosis based on the test result must be supported by a comprehensive judgment of the concerned physician including clinical symptoms and other relevant test results.

MATERIALS SUPPLIED

REF CFPC-114

Components of **ichroma™ COVID-19 Ab**

- Cartridge Box:
 - Cartridge 25
 - Detector 25
 - Detector diluent 1
 - ID chip 1
 - Instruction for use 1

MATERIALS REQUIRED BUT SUPPLIED ON DEMAND

Following items can be purchased separately from **ichroma™ COVID-19 Ab**.

Please contact our sales division for more information.

- **ichroma™ II** **REF** FPRR021
- **ichroma™ M2** **REF** FPRR031

SAMPLE COLLECTION AND PROCESSING

The sample type for **ichroma™ COVID-19 Ab** is human whole blood/serum/plasma.

- It is recommended to test the sample within 24 hours after collection.
- The serum or plasma should be separated from the clot by centrifugation within 3 hours after the collection of whole blood. If longer storage is required, e.g. if the test could not be performed within 24 hours, serum or plasma should be immediately frozen below -20 °C. The freezing storage of sample up to 3 months does not affect the quality of results.
- However, the whole blood sample should not be kept in a freezer in any case.
- Once the sample was frozen, it should be thawed one time and only for test, because repeated freezing and thawing can result in the changed test values.

TEST SETUP

- Check the contents of **ichroma™ COVID-19 Ab**: Sealed cartridges, detectors, detector diluent, ID chip and instruction for use.
- Ensure that the lot number of the cartridge matches that of the detector tubes, detector diluent as well as an ID Chip.
- If the sealed cartridge, the detector and the detector diluent have been stored in a refrigerator, place them on a clean and flat surface at room temperature for at least 30 minutes before testing.
- Avoid directly windy place. The air flow can affect the flow of samples.
- Turn on the instrument for ichroma™ test.
(Please refer to the instrument operation manual for **ichroma™ COVID-19 Ab** test for the concrete information and instructions).

TEST PROCEDURE

■ ichroma™ II

< Multi test mode >

- 1) Transfer 150 µL of detector diluent using a pipette to detector tube containing a granule. When the granule form is completely dissolved in the tube, it becomes detection buffer.
- 2) Draw 10 µL of sample (Human whole blood/serum/plasma/control) with a pipette, and add into detector tube(③) immediately.
- 3) Close the lid of the detector tube and shake about 10 times or more until mix well. The mixture must be used immediately.
- 4) Pipette out 75µL of a sample mixture and load it into the sample well on the cartridge.
- 5) Leave the cartridge at room temperature for 10 minutes before inserting the device into the holder.
⚠ Scan the sample loaded cartridge immediately when the incubation time is over. If not, it will cause inaccurate test result.
- 6) To scan the sample-loaded cartridge, insert it into the cartridge holder of the instrument for ichroma™ tests. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.
- 7) Tap the “START” button on the instrument for ichroma™ test to start the scanning process.
- 8) The instrument for ichroma™ tests will start scanning the sample loaded cartridge immediately.
- 9) Read the test result on the display screen of the instrument for ichroma™ tests.

< Single test mode >

- 1) The test procedure is same with “Multi test mode”.
(Multi Test mode ① - ④)
- 2) Insert sample mixture loaded cartridge into the cartridge holder immediately of the instrument for ichroma™ tests. Ensure proper orientation of the

cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.

- 3) Tap the “START” button on the instrument for ichroma™ test.
- 4) Cartridge goes inside the instrument for ichroma™ tests and will automatically start scanning the cartridge after 10 minutes.
- 5) Read the test result on the display screen of the instrument for ichroma™ tests.

■ ichroma™ M2

< Read Now mode >

- 1) The test procedure is same with “ichroma™ II Multi test mode ① - ④”.
- 2) Leave the cartridge at room temperature for 10 minutes before inserting the device into the cartridge holder of ichroma™ M2.
⚠ Scan the sample loaded cartridge immediately when the incubation time is over. If not, it will cause inexact test result.
- 3) To scan the sample-loaded cartridge, insert it into the cartridge holder of the instrument for ichroma™ tests. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.
- 4) The instrument will automatically start scanning the cartridge. Do not remove the cartridge or touch the reader during scanning.
- 5) Read the test result on the display screen of the instrument.
- 6) When the cartridge is removed from cartridge holder, the display will show “Read Now” as a standby state.

< Walk Away mode >

- 1) Check the display “Walk Away” on the ichroma™ M2 screen.
- 2) The test procedure is same with “ichroma™ II Multi test mode ① - ④”.
- 3) After loading sample mixture, press the key button on ichroma™ M2 instrument to start the test.
⚠ Do not insert the cartridge in this stage. If the cartridge is inserted in this stage, it will be occurred “error”.
- 5) After 5 seconds, “insert cartridge” will displayed on the screen with beep sounds. At this time, insert the mixture loaded cartridge into the holder. Ensure proper orientation of the cartridge before pushing it all the way inside the cartridge holder. An arrow is marked on the cartridge especially for this purpose.
⚠ Be sure to insert the cartridge within 10 seconds, if not, it will be occurred “error”.
- 6) The instrument will automatically start scanning the cartridge after reaction time.
⚠ When the cartridge is inserted, reaction time is displayed. The reaction time will be counted down from step “④”, so initial start time is not 10 minutes. This is normal condition.
- 7) Read the test result on the display screen of the instrument.

- 8) When the cartridge is removed from cartridge holder, the display will show "Walk Away" as a standby state.

INTERPRETATION OF TEST RESULT

- The instrument for **ichroma™** tests calculates the test result automatically and displays 'Positive' / 'Negative' / 'Indeterminate' with ancillary value, cut-off index (COI).

| Cut-off index (COI) | Result | Note |
|---------------------|------------------------|---------------------------|
| < 0.9 | Negative for IgG / IgM | No need to retest |
| 0.9 ≤ Titer < 1.1 | Indeterminate | Need to retest |
| ≥ 1.1 | Positive for IgG / IgM | Need to confirmation test |

- If the test result is "Negative" even though the patient has significant infectious symptoms, it should be recommended to conduct additional test including PCR or culture test.
- The accurate determination of test result as "Positive" should be confirmed by additional clinical evaluation.
- "Negative" result should be considered with possibilities of other infections. Positive result should be considered with additional infections by another pathogenic bacterium.

PERFORMANCE CHARACTERISTICS

Analytical Sensitivity

Cut-off

The **ichroma™ COVID-19 Ab** test result indicates 'positive' or 'negative' of a sample defined by the algorithm of **ichroma™** reader based on COI (cut-off index).

| Cut-off index (COI) | Result |
|---------------------|------------------------|
| < 0.9 | Negative for IgG / IgM |
| 0.9 ≤ Titer < 1.1 | Indeterminate |
| ≥ 1.1 | Positive for IgG / IgM |

Analytical specificity

Cross-reactivity

Biomolecules such as below the ones in the table were added to the test sample(s) at concentrations much higher than their normal physiological levels in the blood. **ichroma™ COVID-19 Ab** test results did not show any significant cross-reactivity with these biomolecules.

| Name | Sample type |
|--------------------------------|--|
| Cytomegalovirus(CMV) | Positive serum |
| Epstein-Barr virus(EBV) | Positive serum |
| Hepatitis A virus(HAV) | Positive serum |
| Hepatitis C virus(HCV) | Positive serum |
| Hepatitis B virus(HBV) | Positive serum |
| Herpes simplex virus(HSV) | Positive serum |
| Rubella virus | Positive serum |
| Varicella-zoster virus(VZV) | Positive serum |
| Treponema pallidum | Positive serum |
| Anti Nuclear Antibody(ANA) | Positive serum |
| Rheumatoid factor(RF) | Positive serum |
| Early stage of pregnancy | Pregnant women sample |
| Middle stage of pregnancy | Pregnant women sample |
| Hepatitis B antibody(anti-HBs) | Hepatitis B (HBsAg) Ab positive sample |

| | |
|-----------------------|----------------|
| Influenza A | Positive serum |
| Influenza B | Positive serum |
| RSV | Positive serum |
| Mycoplasma pneumoniae | Positive serum |

Interference

Interference materials such as below the ones in the table were added to the test sample(s) the same as the below concentrations. **ichroma™ COVID-19 Ab** test results did not show any significant interference with these materials.

| Material | Concentration |
|----------------------|---------------------|
| Li-Heparin | 100,000 U/L |
| Na-Heparin | 100,000 U/L |
| Na-EDTA | 1.6 mg/mL (4 μM) |
| K ₂ -EDTA | 1.6 mg/mL (4 μM) |
| Sodium citrate | 25 mg/mL (0.085 μM) |
| Hemoglobin | 2 mg/ml |
| BSA | 60 mg/ml |
| Bilirubin | 0.24 mg/mL (400 μM) |
| Triglycerides | 1.5 mg/ml |
| Cholesterol | 7.7 mg/mL (20 mM) |

Precision

Between lots

One person tested three different lots of **ichroma™ COVID-19 Ab**, ten times at each concentration of the control standard.

Between persons

Three different persons tested one lot of **ichroma™ COVID-19 Ab**, ten times at each concentration of the control standard.

Between days

One person tested one lot of **ichroma™ COVID-19 Ab** during three days, ten times at each concentration of the control standard.

Between sites

One person tested **ichroma™ COVID-19 Ab** at three different site, ten times at each concentration of the control standard.

[IgG result]

| Cal No. | Between lot | | Between person | |
|---------|----------------|---------------|----------------|---------------|
| | Positive / No. | Positive rate | Positive / No. | Positive rate |
| 1 | 0/30 | 0% | 0/30 | 0% |
| 2 | 30/30 | 100% | 30/30 | 100% |
| 3 | 30/30 | 100% | 30/30 | 100% |
| Cal No. | Between day | | Between site | |
| | Positive / No. | Positive rate | Positive / No. | Positive rate |
| 1 | 0/30 | 0% | 0/30 | 0% |
| 2 | 30/30 | 100% | 30/30 | 100% |
| 3 | 30/30 | 100% | 30/30 | 100% |

[IgM result]













| Cal No. | Between lot | | Between person | |
|------------|----------------|---------------|----------------|---------------|
| | Positive / No. | Positive rate | Positive / No. | Positive rate |
| 1 | 0/30 | 0% | 0/30 | 0% |
| 2 | 30/30 | 100% | 30/30 | 100% |
| 3 | 30/30 | 100% | 30/30 | 100% |

| Cal No. | Between day | | Between site | |
|------------|----------------|---------------|----------------|---------------|
| | Positive / No. | Positive rate | Positive / No. | Positive rate |
| 1 | 0/30 | 0% | 0/30 | 0% |
| 2 | 30/30 | 100% | 30/30 | 100% |
| 3 | 30/30 | 100% | 30/30 | 100% |

REFERENCES

- Huang LR *et al.* Evaluation of Antibody Responses Against SARS Coronaviral Nucleocapsid or Spike Proteins by Immunoblotting or ELISA (2004) J Med Virol. 73: 33.
- Woo PC *et al.* Longitudinal profile of immunoglobulin G (IgG), IgM, and IgA antibodies against the (SARS) coronavirus nucleocapsid protein in patients with pneumonia (2004) Clin Diagn Lab Immunol. 11: 665.
- Wu HS *et al.* SARS-Associated Coronavirus Diagnostic kit_ development of an ELISA-based antibody detection test with a cocktail of nucleocapsid and spike SARS-CoV proteins (2008) J Clin Microbiol. 43: 3054.
- Trivedi SU *et al.* Development and Evaluation of a Multiplexed Immunoassay for Simultaneous Detection of Serum IgG Antibodies to Six Human Coronaviruses (2019) Sci Rep. 9: 1390

Note: Please refer to the table below to identify various symbols.

| | |
|---|---|
|  | Sufficient for <n> tests |
|  | Read instruction for use |
|  | Use by Date |
|  | Batch code |
|  | Catalog number |
|  | Caution |
|  | Manufacturer |
|  | Authorized representative of the European Community |
|  | In vitro diagnostic medical device |
|  | Temperature limit |
|  | Do not reuse |
|  | This product fulfills the requirements of the Directive 98/79/EC on in vitro diagnostic medical devices |

For technical assistance; please contact:

Boditech Med Inc.'s Technical Services

Tel: +82 33 243-1400

E-mail: sales@boditech.co.kr

**Boditech Med Incorporated**

43, Geodudanji 1-gil, Dongnae-myeon,

Chuncheon-si, Gang-won-do, 24398

Republic of Korea

Tel: +(82) -33-243-1400

Fax: +(82) -33-243-9373

www.boditech.co.kr

**Obelis s.a**

Bd. Général Wahis 53, 1030 Brussels, BELGIUM

Tel: +(32) -2-732-59-54

Fax: +(32) -2-732-60-03

E-Mail: mail@obelis.net

