

Standard : ASTM D3080, BS 1377-7, AASHTO T236, NF P094 071-1/2

Application: This apparatus is used to test resisting shear strength of soil specimen.

The direct shear apparatus is driven by a stepper motor and controlled by a microprocessor system, with load cell and linear electronic transducers to operate independently to perform standard or cyclic tests, programmable by keyboard input and obtaining all data concerning load, horizontal displacements from the display.

Specification:

- 1.LCD Display
2. Accepts specimens 61.8 x 20 mm.
3. Area of soil sample: 30 cm²
- 4.Level ratio: 1:12
- 5.Vertical Loading: 50, 100, 200, 300, 400(KPa)
- 6.Max. horizontal loading:1.2kN
- 7.Variable speed, 0.01 to 2.99mm./minute
- 8.AC220V +/- 10%



C S T INSTRUMENTS(THAILAND)LTD.,PART.

95 Romklao Road,
Klongsam-Praves,
Ladkrabang, Bangkok
Tel.02-915-1693-5

Dealer