



Shelf-Life Extension Granted for Flowflex™ COVID-19 Antigen Home Test and On/Go™ One COVID-19 Antigen Home Test

November 18, 2022

To Whom It May Concern:

ACON Laboratories has submitted additional real-time stability data to the US FDA for the Flowflex™ COVID-19 Antigen Home Test (EUA210494). Based on that data, the FDA has granted a twenty one (21) month shelf-life for both the Flowflex COVID-19 Antigen Home Test and the On/Go™ One COVID-19 Antigen Home Test when stored at 2°C - 30°C.

This shelf-life extension represents a nine (9) month extension of the expiration date indicated on the product. Please refer to the table below for the new expiration dates of On/Go One COVID-19 Antigen Home Tests.

Lot Numbers	Printed Expiration Date (12 mo)	New Expiration Date (21 mo)
COV2010036 COV2010083 COV2010088	January 2023	October 2023
COV2020004 COV2020011 COV2020014 COV2020020 COV2020026 COV2020061 COV2020077	February 2023	November 2023

Best regards,

Michael Lynch,
VP, Sales and Marketing