

Test Certificate

concerning the test of influences on MRI systems caused by
non-Siemens products

1. Test certificate, product designation and manufacturer

This is to certify that the following product

NordicNeurolab Monitor

Tested with Serial No. 101121 LCD 3.1.3

of the manufacturer:

NordicNeurolab AS

Møllendalsveien 1

N-5009 Bergen

does not impair the functioning of the MRI systems of Siemens Shenzhen Magnetic Resonance Ltd. specified in section 2.

All tests were performed by SHS Magnetic Resonance in Erlangen, we confirm that Siemens Shenzhen Magnetic Resonance Ltd. is the legal manufacturer.

Potential functional restrictions of the MRI systems are indicated in section 4.

Possible adverse effects of the MRI systems on the product indicated above are expressly not subject of the test on which this certificate is based. Therefore, this certificate does not imply non-interference of the above-mentioned product through the MRI system.

2. 0.55T MAGNETOM systems involved

Seq. no. MR-System or Option

1 Magnetom free.Max

2 Magnetom free.Star

3. Applications

The “**NNL Monitor**” is intended for use on MR Systems till a Magnet field intensity of 0.55 Tesla:

- MRI Imaging

4. Limitations

During the use of this product the following limitations of functionality and/or applications of the MRI systems, options and accessories apply:

- Warning Label according to user manual has to be added to the power supply.

5. Warnings

The following safety precautions must be observed when using this **NNL Monitor** in connection with the MRI systems listed in section 2:

- N.a.

6. Validity

This test certificate shall be valid until revoked by Siemens Healthcare GmbH or Siemens Shenzhen Magnetic Resonance Ltd.

Siemens Shenzhen Magnetic Resonance Ltd.



Electronically signed by:
Hai Ting Li
Reason: I am approving
this document
Date: May 13, 2022 10:11
GMT+8

Hai Ting Li Head of SSMR QT

Siemens Healthcare GmbH



Electronically signed
by: Joerg Teiche
Reason: I.V.
Date: May 12, 2022
15:50 GMT+2

Joerg Teiche Vice President Quality MR



Electronically signed by:
Peter Gall
Reason: Document
Execution
Date: May 13, 2022 08:18
GMT+1

Dr. Peter Gall Vice President Portfolio and Partners MR