



TESTING KITS FOR DRUGS

Procedure Card for IDT0350





FLUNITRAZEPAM/ BENZODIAZEPINES IDENTIFICATION





Substance Test



Read Before Use

As a precaution, gloves and eye protection should be worn when handling drug substances and using test units. Do not puncture any part of unit. Do not remove probe once a capsule is broken. In the unlikely event of a spill, rinse affected area thoroughly with water. Keep out of reach of children. All results are presumptive. If necessary, further testing by a competent drug testing lab can provide additional identification.

Test Preparation	1	Remove the Smart-Tip™ probe from test unit.	2	Sample suspected substance.	
				Powder	Tablet
				Touch Smart-Tip™ probe to sample.	Firmly press Smart-Tip™ probe on the tablet while scrubbing in a circular motion until the tip is completely covered with substance.
					
Test Preparation	3	Lock Smart-Tip™ probe into test unit. Press firmly.	4	5	6
					
Identification	7	8	9	10	11

Results		The appearance of a red-purple (wine) color indicates the presence of Flunitrazepam .
		The appearance of a yellow color indicates the possible presence of Nitrazepam, Clonazepam or Nimetazepam .
		The appearance of a green color indicates the possible presence of Lorazepam or Temazepam .
		If none of the above colors develop, the above listed substances are <u>not</u> present.



Disposal

The test kit is safe to dispose of in the trash.

01-19



Manufactured by IDenta Ltd
www.identa-corp.com
Patented Technology
Made in Israel